



**HERMON LABORATORIES**



**Certificate of Conformity**  
with EN/IEC 60601-1:1990; UL 60601-1:2003,  
EN/IEC 60601-2-38:1996

**N° VITSAF\_18769C**

Product definition..... : Hospital bed  
Model..... : Vitalgo Total Lift Bed HS/NU/HM  
Ratings and principal characteristics : 120 VAC, 60 Hz, 5A max or 230 VAC, 50 Hz, 3A max  
Applicant/Manufacturer ..... : VitalGo Systems Ltd.  
Address ..... : 7 Florinis St., Nicosia, Cyprus  
..... (45 Pinchas Rozen St., Tel Aviv 69512, Israel)

This Certificate of Conformity has been granted to the applicant based on the results of tests and measurements, performed by Hermon Laboratories on a sample of the above-mentioned product from Sept 2008.

The tested product satisfies the requirements of:  
EN/IEC 60601-1:1990; UL 60601-1:2003, EN/IEC 60601-2-38:1996.

Measurements/tests results are contained in the test report:  
VITSAF\_EN.18796

**Yigal Cohen,**  
**Product Safety, Group manager**  
**Hermon Laboratories Ltd.**



Sep 24, 2008

The comments in the associated (if applicable) test report/s shall be taken into account and used in conjunction with this certificate. The "A2LA Accredited" symbol endorsement applies only to the tests and calibrations that are listed in the scope of Hermon Laboratories accreditation.